

ACCURATE HEMOGLOBIN AND HEMATOCRIT RESULTS WITH ONE SIMPLE TEST



2 results in 1 test

Easy to use

No calibration required

Data management options available



EKF
DIAGNOSTICS

SCIENCE MADE SIMPLE



EASY TO USE

- User friendly features minimises training time
- Step by step instructions on screen
- Backlit touch screen
- User selectable language menu
- Soft-load cuvette holder minimizes risk of contamination

PRACTICAL AND PORTABLE

- Hemoglobin and hematocrit results from one sample available after 25–60 seconds (depending on the concentration)
- Venous, arterial or capillary blood
- Sample volume only 8 or 10 µl depending on cuvette type
- nxt shape microcuvette minimises air bubbles
- Compact in size and weighing just 700g
- No maintenance required. Auto self-test
- Integrated rechargeable battery (100 hours)


ACCURATE AND RELIABLE

- Operating ambient temperature 15°C - 40°C
- Photometric azide methemoglobin method
- Measuring range: 0–25.6 g/dL; 0–15.9 mmol/L
- Imprecision: CV <2 %
- Linearity: 0–20 g/dl: ±0.3 g/dl; >20 g/dl: ±0.7 g/dl
- Calibrated to NCCLS reference method
- Stores up to 4,000 patient results
- Control cuvette included

HEMO CONTROL MANAGER

- Bar code identification of patients, operators and control levels
- Connect to PC via serial port or Bluetooth
- Quality Control lockout function
- Stores 4,000 patient results and 500 QC results

Simple, accurate and reliable Hemoglobin and Hematocrit measurement in 3 easy steps



1

Collect blood sample.



2

Put microcuvette into analyser.



3

Result appears in 25–60 seconds.

Distributed by

Manufacturer

EKF diagnostic GmbH
Ebendorfer Chaussee 3
39179 Barleben, Germany
Tel. +49 (0)39203 7850



Approved for the European Market

Sales

EKF Diagnostics Holdings plc
Avon House, 19 Stanwell Road,
Penarth, Cardiff, UK, CF64 2EZ
Tel. +44 (0)29 20 710570

www.ekfdiagnostics.com



SCIENCE MADE SIMPLE

Disclaimer: EKF-diagnostic GmbH reserves the right to change the product specifications without prior notice.

Revision 3.0-11/2013