K-ASSAY® *The Assay You Can Trust . . .*

Immunoassay Reagents
for Chemistry Analyzers

CATALOG

May 2025

KAMIYA BIOMEDICAL COMPANY

800-526-4925 / 206-575-8068 • FAX: 206-575-8094
www.k-assay.com


diagnostics
antibodies
biochemicals
K

42
Years
In Business Since
1983
KAMIYA BIOMEDICAL COMPANY

K-ASSAY® *The Assay You Can Trust . . .*

Immunoassay Reagents for Chemistry Analyzers

Advantages

- ☞ Adaptable to most chemistry analyzers
- ☞ Liquid-stable reagents: no dissolving, mixing, or dilution
- ☞ No predilution of samples
- ☞ Wide measuring range
- ☞ Most products have 1-month stability after opening, ≥ 1-year expiration from date of manufacture

Products

Anemia	Cat. No.
Ferritin	KAI-095
Transferrin	KAI-023
UIBC	KAI-300

Cardiac Assessment	Cat. No.
CRP, High-Sensitive	KAI-160
CRP, Wide Range	KAI-082
D-Dimer	KAI-090
D-Dimer, High-Sensitive**	KAI-102**
Ferritin	KAI-095
Fibrinogen (Antigen)	KAI-135
Microalbumin	KAI-019

Coagulation	Cat. No.
D-Dimer	KAI-090
D-Dimer, High-Sensitive**	KAI-102**
Factor XIII	KAI-105
Plasma FDP*	KAI-111*
Serum FDP*	KAI-320*
Urine FDP*	KAI-325*
Fibrinogen (Antigen)	KAI-135

Diabetes/Stroke	Cat. No.
Alpha-1 Microglobulin	KAI-056
Cystatin C	KAI-074
Hemoglobin A1c	KAI-196
Insulin	KAI-071
Microalbumin	KAI-019

Infectious Diseases	Cat. No.
Beta-2 Microglobulin	KAI-280
Helicobacter pylori	KAI-240
Krebs von den Lungen-6 [KL-6 (Ver.2)]	KAI-250
Remnant Lipoprotein Cholesterol (RLP-C)	KAI-260

Inflammation	Cat. No.
Anti-Streptolysin O	KAI-078
Complement C3	KAI-009
Complement C4	KAI-010
C-Reactive Protein (CRP)	KAI-026
CRP, High-Sensitive	KAI-160
CRP, Wide Range	KAI-082
Fibrinogen (Antigen)	KAI-135
Rheumatoid Factor (Ver.2)	KAI-230

Kidney Function	Cat. No.
Cystatin C	KAI-074
Microalbumin	KAI-019

Lipid Assessment	Cat. No.
Apolipoprotein AI	KAI-002
Apolipoprotein AII*	KAI-003*
Apolipoprotein B	KAI-004
Apolipoprotein CII*	KAI-005*
Apolipoprotein CIII*	KAI-006*
Apolipoprotein E*	KAI-007*
Lipoprotein(a)	KAI-044

Nutritional Assessment	Cat. No.
Ferritin	KAI-095
Prealbumin	KAI-053
Transferrin	KAI-023
UIBC	KAI-300

Serum Proteins	Cat. No.
Alpha-1 Microglobulin	KAI-056
Haptoglobin	KAI-022
IgA	KAI-013
IgG	KAI-014
IgM	KAI-015
Transferrin	KAI-023

* For research use only in the U.S. Not for use in diagnostic procedures in the U.S.
** Not for use in the U.S.

Immunoassay Reagents for Chemistry Analyzers

Alpha-1 Microglobulin

Method: Immunoturbidimetric
Format: Liquid
Linearity: 0.2 - 34.0 mg/L (urine)
1.0 - 137 mg/L (serum/plasma)
Sample: Urine or Serum/Plasma

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-056	**Alpha-1 Microglobulin** R1: 1 x 50 mL buffer R2: 1 x 50 mL antibody Tests / Kit: 357 serum/plasma (H917) 500 urine (H917)

Required Calibrator (Choose One):

KAI-067C **Urine Alpha-1 Microglobulin Calibrator**
4 x 1 mL, liquid
For 5 point calibration

OR

KAI-068C **Serum/Plasma Alpha-1 Microglobulin Calibrator**
5 x 1 mL, liquid
For 6 point calibration

Anti-Streptolysin O

Method: Immunoturbidimetric
Format: Liquid
Linearity: 20 - 1,000 IU/mL
Sample: Serum

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-078	**ASO** R1: 4 x 20 mL buffer R2: 4 x 20 mL Streptolysin O Tests / Kit: 571 (H917)

Required Calibrator:

KAI-079C **ASO Calibrator**
5 x 1 mL, liquid
For 5 point calibration

Recommended Control:

K55C-4M **ASO/RF/CRP Control**
L1: 2 x 2 mL, lyophilized
L2: 2 x 2 mL, lyophilized

Apolipoprotein AI

Method: Immunoturbidimetric
Format: Liquid
Linearity: 20 - 300 mg/dL
Sample: Serum

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-002	**Apo AI** R1: 3 x 20 mL buffer R2: 1 x 20 mL antibody Tests / Kit: 285 (H917)

Required Calibrator:

KAI-008C **Apo AI/B Calibrator**
1 x 1 mL, lyophilized
2 x 2 mL, liquid diluent
For 6 point calibration

Recommended Control:

K310C-4M **Apo AI / B Control**
L1: 2 x 1 mL, lyophilized
L2: 2 x 1 mL, lyophilized

Immunoassay Reagents for Chemistry Analyzers

Apolipoprotein AII

Method: Immunoturbidimetric
Format: Liquid
Linearity: 10 - 100 mg/dL
Sample: Serum

For Research Use Only in the U.S. Not for use in diagnostic procedures in the U.S.

Cat. No.	Description
KAI-003*	**Apo AII***
	R1: 4 x 18 mL buffer
	R2: 1 x 10 mL antibody
	Tests / Kit: 285 (H917)

Required Calibrator:

KAI-041C*	**Apo AII/CII/CIII Calibrator***
	1 x 1 mL, lyophilized
	2 x 2 mL, liquid diluent
	For 6 point calibration

Recommended Control:

K315C-4M*	**Apo AII / CII / CIII / E Control***
	L1: 2 x 1 mL, lyophilized
	L2: 2 x 1 mL, lyophilized

Apolipoprotein B

Method: Immunoturbidimetric
Format: Liquid
Linearity: 25 - 250 mg/dL
Sample: Serum

For *In Vitro* Diagnostic Use.



Cat. No.	Description
KAI-004	**Apo B**
	R1: 3 x 20 mL buffer
	R2: 1 x 20 mL antibody
	Tests / Kit: 285 (H917)

Required Calibrator:

KAI-008C	**Apo AI/B Calibrator**
	1 x 1 mL, lyophilized
	2 x 2 mL liquid diluent
	For 6 point calibration

Recommended Control:

K310C-4M	**Apo AI / B Control**
	L1: 2 x 1 mL, lyophilized
	L2: 2 x 1 mL, lyophilized

Apolipoprotein CII

Method: Immunoturbidimetric
Format: Liquid
Linearity: 1 - 15 mg/dL
Sample: Serum

For Research Use Only in the U.S. Not for use in diagnostic procedures in the U.S.

Cat. No.	Description
KAI-005*	**Apo CII***
	R1: 3 x 20 mL buffer
	R2: 1 x 20 mL antibody
	Tests / Kit: 285 (H917)

Required Calibrator:

KAI-041C*	**Apo AII/CII/CIII Calibrator***
	1 x 1 mL, lyophilized
	2 x 2 mL liquid diluent
	For 6 point calibration

Recommended Control:

K315C-4M*	**Apo AII / CII / CIII / E Control***
	L1: 2 x 1 mL, lyophilized
	L2: 2 x 1 mL, lyophilized

* For research use only in the U.S. Not for use in diagnostic procedures in the U.S.

Immunoassay Reagents for Chemistry Analyzers

Apolipoprotein CIII

Method: Immunoturbidimetric
Format: Liquid
Linearity: 3 - 30 mg/dL
Sample: Serum

For Research Use Only in the U.S. Not for use in diagnostic procedures in the U.S.

Cat. No.	Description
KAI-006*	**Apo CIII***
	R1: 3 x 20 mL buffer
	R2: 1 x 20 mL antibody
	Tests / Kit: 285 (H917)

Required Calibrator:

KAI-041C*	**Apo AII/CII/CIII Calibrator***
	1 x 1 mL, lyophilized
	2 x 2 mL liquid diluent
	For 6 point calibration

Recommended Control:

K315C-4M*	**Apo AII / CII / CIII / E Control***
	L1: 2 x 1 mL, lyophilized
	L2: 2 x 1 mL, lyophilized

Apolipoprotein E

Method: Immunoturbidimetric
Format: Liquid
Linearity: 1 - 15 mg/dL
Sample: Serum

For Research Use Only in the U.S. Not for use in diagnostic procedures in the U.S.

Cat. No.	Description
KAI-007*	**Apo E***
	R1: 3 x 20 mL buffer
	R2: 1 x 20 mL antibody
	Tests / Kit: 285 (H917)

Required Calibrator:

KAI-025C*	**Apo E Calibrator***
	1 x 1 mL, lyophilized
	2 x 2 mL liquid diluent
	For 6 point calibration

Recommended Control:

K315C-4M*	**Apo AII / CII / CIII / E Control***
	L1: 2 x 1 mL, lyophilized
	L2: 2 x 1 mL, lyophilized

Beta-2 Microglobulin

Method: Immunoturbidimetric
Format: Liquid
Linearity:
Serum/Plasma: 0.2 - 80.0 mg/L
Urine: 30 - 8,000 µg/L or
0.03 - 8.00 mg/L
Sample: Serum, Plasma, Urine

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-280	**Beta-2 Microglobulin**
	R1: 1 x 45 mL buffer
	R2: 1 x 20 mL antibody
	Tests / Kit:
	250 serum / plasma (H917)
	333 urine (H917)

Required Calibrator (Choose One):

KAI-281C	**B2MG Serum / Plasma Calib.**
	5 x 2 mL, liquid
	For 6 point calibration
OR	
KAI-282C	**B2MG Urine Calibrator**
	5 x 2 mL, liquid
	For 6 point calibration

Recommended Control Options:

K283C-4M	**B2MG / Ferritin - Ser / Pl Control**
	L1: 2 x 3 mL, liquid
	L2: 2 x 3 mL, liquid
OR	
K284C-2M	**B2MG Urine Control**
	L1: 1 x 2 mL, liquid
	L2: 1 x 2 mL, liquid

* For research use only in the U.S. Not for use in diagnostic procedures in the U.S.

800-526-4925 / 206-575-8068 • FAX: 206-575-8094
www.k-assay.com

Immunoassay Reagents for Chemistry Analyzers

Complement C3

Method: Immunoturbidimetric
Format: Liquid
Linearity: 30 - 350 mg/dL
Sample: Serum

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-009	**Complement C3** R1: 4 x 20 mL buffer R2: 2 x 10 mL antibody Tests / Kit: 363 (H917)

Required Calibrator:

KAI-016C	**Multi-Analyte Calibrator** 6 x 1 mL, liquid For 6 point calibration

Recommended Control:

K49C-4M	**Liquid ITA Control** L1: 2 x 2 mL, liquid L2: 2 x 2 mL, liquid

Complement C4

Method: Immunoturbidimetric
Format: Liquid
Linearity: 8 - 80 mg/dL
Sample: Serum

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-010	**Complement C4** R1: 4 x 20 mL buffer R2: 2 x 10 mL antibody Tests / Kit: 363 (H917)

Required Calibrator:

KAI-016C	**Multi-Analyte Calibrator** 6 x 1 mL, liquid For 6 point calibration

Recommended Control:

K49C-4M	**Liquid ITA Control** L1: 2 x 2 mL, liquid L2: 2 x 2 mL, liquid

C-Reactive Protein

Method: Immunoturbidimetric
Format: Liquid
Linearity: 1 - 250 mg/L
Sample: Serum

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-026	**CRP** R1: 4 x 20 mL buffer R2: 2 x 10 mL antibody Tests / Kit: 333 (H917)

Required Calibrator:

KAI-012C	**CRP Calibrator** 6 x 1 mL, liquid For 6 point calibration

Recommended Control Options:

K49C-4M	**Liquid ITA Control** L1: 2 x 2 mL, liquid L2: 2 x 2 mL, liquid
OR	
K55C-4M	**ASO/RF/CRP Control** L1: 2 x 2 mL, lyophilized L2: 2 x 2 mL, lyophilized

Immunoassay Reagents for Chemistry Analyzers

C-Reactive Protein, High-Sensitive

Method: Immunoturbidimetric
Format: Liquid
Linearity: 0.20 - 20.0 mg/L
Sample: Serum, Plasma

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-160	**hsCRP** R1: 1 x 50 mL buffer R2: 1 x 10 mL antibody Tests / Kit: 200 (H917)

Required Calibrator:

KAI-161C	**hsCRP Calibrator** 4 x 1 mL, liquid For 5 point calibration

Recommended Control:

K180C-2M	**hsCRP Control** L1: 1 x 3 mL, liquid L2: 1 x 3 mL, liquid

C-Reactive Protein, Wide Range

Method: Immunoturbidimetric
Format: Liquid
Linearity: 0.05 - 160 mg/L
Sample: Serum, Plasma

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-082	**CRP (3)** R1: 1 x 50 mL buffer R2: 1 x 50 mL antibody Tests / Kit: 333 (H917)

Required Calibrator:

KAI-084C	**CRP (3) Calibrator E** 5 x 2 mL, liquid For 6 point calibration

Recommended Control Options:

K180C-2M	**hsCRP Control** L1: 1 x 3 mL, liquid L2: 1 x 3 mL, liquid
OR	
K55C-4M	**ASO/RF/CRP Control** L1: 2 x 2 mL, lyophilized L2: 2 x 2 mL, lyophilized

Cystatin C

Method: Immunoturbidimetric
Format: Liquid
Linearity: 0.06 - 8.00 mg/L
Sample: Serum, Plasma

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-074	**Cystatin C** R1: 2 x 10 mL buffer R2: 2 x 10 mL antibody Tests / Kit: 200 (H917)

Required Calibrator:

KAI-099C	**Cystatin C Calibrator** 6 x 1 mL, liquid For 6 point calibration

Recommended Control:

K110C-4M	**Cystatin C Control** L1: 2 x 1 mL, liquid L2: 2 x 1 mL, liquid

Immunoassay Reagents for Chemistry Analyzers

D-Dimer

Method: Immunoturbidimetric
Format: Liquid
Linearity: 0.5 - 30 µg/mL
Sample: Serum, Plasma

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-090	**D-Dimer** R1: 1 x 16 mL buffer R2: 1 x 8.5 mL antibody Tests / Kit: 114 (H917)

Required Calibrator:

KAI-091C	**D-Dimer Calibrator** 6 x 1 mL, lyophilized For 6 point calibration

Recommended Control:

K92C-4M	**D-Dimer Control** L1: 2 x 1 mL, lyophilized L2: 2 x 1 mL, lyophilized

D-Dimer, High-Sensitive

Method: Immunoturbidimetric
Format: Liquid
Linearity: 0.03 - 8.00 µg/mL or
30 - 8,000 ng/mL
Sample: Serum, Plasma

Not For Use in the U.S.

Cat. No.	Description
KAI-102**	**D-Dimer, High-Sensitive**** R1: 2 x 18 mL buffer R2: 1 x 7 mL antibody Tests / Kit: 175 (H917)

Required Calibrator:

KAI-103C**	**High-Sensitive D-Dimer Calibrator**** 5 x lyophilized stock 1 x 40 mL liquid diluent For 6 point calibration

Recommended Control:

K115C-10M**	**High-Sensitive D-Dimer Control**** L1: 5 x 0.5 mL, lyophilized L2: 5 x 0.5 mL, lyophilized

Factor XIII

Method: Immunoturbidimetric
Format: Liquid
Linearity: 2.3 - 140 %
Sample: Plasma

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-105	**Factor XIII** R1: 2 x 9.5 mL buffer R2: 1 x 6 mL antibody Tests / Kit: 120 (H917)

Required Calibrator:

KAI-106C	**Factor XIII Calibrator** 5 x 1.0 mL, lyophilized 1 x 40 mL liquid diluent For 6 point calibration

Recommended Control:

K135C-10M	**Factor XIII Control** L1: 5 x 0.5 mL, lyophilized L2: 5 x 0.5 mL, lyophilized

** Not for use in the U.S.

Immunoassay Reagents for Chemistry Analyzers

Plasma FDP

Method: Immunoturbidimetric
Format: Liquid
Linearity: 2 - 80 µg/mL
Sample: Plasma

For Research Use Only in the U.S. Not for use in diagnostic procedures in the U.S.

Cat. No.	Description
KAI-111*	**Plasma FDP***
	R1: 2 x 10.5 mL buffer
	R2: 1 x 6 mL antibody
	Tests / Kit: 120 (H917)

Required Calibrator:

KAI-112C*	**Plasma FDP Calibrator***
	5 x 1.0 mL, lyophilized
	1 x 40 mL liquid diluent
	For 6 point calibration

Recommended Control:

K131C-10M*	**Plasma FDP Control***
	L1: 5 x 0.5 mL, lyophilized
	L2: 5 x 0.5 mL, lyophilized

Serum FDP

Method: Immunoturbidimetric
Format: Liquid
Linearity: 0.2 - 80 µg/mL
Sample: Serum

For Research Use Only in the U.S. Not for use in diagnostic procedures in the U.S.

Cat. No.	Description
KAI-320*	**Serum FDP***
	R1: 2 x 12 mL buffer
	R2: 1 x 5 mL antibody
	Tests / Kit: 125 (H917)

Required Calibrator:

KAI-321C*	**Serum FDP Calibrator***
	5 x 1.0 mL, lyophilized
	1 x 40 mL liquid diluent
	For 6 point calibration

Recommended Control Options:

K322C-10M*	**Serum FDP Control***
	L1: 5 x 0.5 mL, lyophilized
	L2: 5 x 0.5 mL, lyophilized

Urine FDP

Method: Immunoturbidimetric
Format: Liquid
Linearity: 0.02 - 3.2 µg/mL
Sample: Urine

For Research Use Only in the U.S. Not for use in diagnostic procedures in the U.S.

Cat. No.	Description
KAI-325*	**Urine FDP***
	R1: 2 x 12 mL buffer
	R2: 1 x 5 mL antibody
	Tests / Kit: 138 (H917)

Required Calibrator:

KAI-326C*	**Urine FDP Calibrator***
	5 x 1.0 mL, lyophilized
	1 x 40 mL liquid diluent
	For 6 point calibration

Recommended Control Options:

K327C-10M*	**Urine FDP Control***
	L1: 5 x 0.5 mL, lyophilized
	L2: 5 x 0.5 mL, lyophilized

Optional Diluent:

KAI-109D*	**Urine FDP Sample Diluent***
	1 x 40 mL liquid diluent
	For dilution of high concentration samples

* For research use only in the U.S. Not for use in diagnostic procedures in the U.S.

Immunoassay Reagents for Chemistry Analyzers

Ferritin

Method: Immunoturbidimetric
Format: Liquid
Linearity: 2.0 - 1,000 ng/mL
Sample: Serum, Plasma

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-095	**Ferritin** R1: 1 x 18 mL buffer R2: 1 x 9 mL antibody Tests / Kit: 128 (H917)

Required Calibrator:

KAI-094C	**Ferritin Calibrator** 4 x 1 mL, liquid For 5 point calibration

Recommended Control:

K283C-4M	**B2MG / Ferritin - Ser / Pl Control** L1: 2 x 3 mL, liquid L2: 2 x 3 mL, liquid

Fibrinogen (Antigen)

Method: Immunoturbidimetric
Format: Liquid
Linearity: 100 - 900 mg/dL
Sample: Plasma

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-135	**Fibrinogen** R1: 2 x 20 mL buffer R2: 1 x 8 mL antibody Tests / Kit: 160 (H917)

Required Calibrator:

KAI-136C	**Fibrinogen Calibrator** 1 x 1 mL, lyophilized For 4 point calibration

Recommended Control:

K63C-10M	**Fibrinogen Control** L1: 5 x 1 mL, lyophilized L2: 5 x 1 mL, lyophilized

H. pylori

Method: Immunoturbidimetric
Format: Liquid
Linearity: 3 - 100 U/mL
Sample: Serum, Plasma

For Research Use Only in the U.S. Not for use in diagnostic procedures in the U.S.

Cat. No.	Description
KAI-240*	**H. pylori*** R1: 1 x 60 mL buffer R2: 1 x 12 mL antibody Tests / Kit: 300 (H917)

Required Calibrator:

KAI-241C*	**H. Pylori Calibrator*** 5 x 2 mL, liquid For 6 point calibration

Recommended Control:

K242C-2M*	**H. pylori Control*** L1: 1 x 2 mL, liquid L2: 1 x 2 mL, liquid

* For research use only in the U.S. Not for use in diagnostic procedures in the U.S.

Immunoassay Reagents for Chemistry Analyzers

Haptoglobin

Method: Immunoturbidimetric
Format: Liquid
Linearity: 20 - 220 mg/dL
Sample: Serum

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-022	**Haptoglobin** R1: 3 x 18 mL buffer R2: 2 x 7 mL antibody Tests / Kit: 254 (H917)

Required Calibrator:

KAI-016C	**Multi-Analyte Calibrator** 6 x 1 mL, liquid For 6 point calibration

Recommended Control:

K49C-4M	**Liquid ITA Control** L1: 2 x 2 mL, liquid L2: 2 x 2 mL, liquid

Hemoglobin A1c

Method: Immunoturbidimetric
Format: Liquid
Linearity: 2.0 - 16.0 %
Sample: Whole Blood

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-196	**Hemoglobin A1c** R1: 1 x 30 mL buffer R2: 1 x 10 mL antibody Hemolysis Reagent: 1 x 125 mL Tests / Kit: 200 (H917)

Cat. No.	Description
KAI-197	**Hemoglobin A1c (L)** R1: 1 x 90 mL buffer R2: 1 x 30 mL antibody Hemolysis Reagent: 1 x 500 mL Tests / Kit: 600 (H917)

Required Calibrator:

KAI-098C	**Hemoglobin A1c Calibrator** 4 x 0.5 mL, lyophilized For 5 point calibration

Recommended Control:

K29C-4M	**Hemoglobin A1c Control** L1: 2 x 0.5 mL, lyophilized L2: 2 x 0.5 mL, lyophilized

Immunoglobulin A

Method: Immunoturbidimetric
Format: Liquid
Linearity: 20 - 800 mg/dL
Sample: Serum

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-013	**IgA** R1: 4 x 20 mL buffer R2: 2 x 10 mL antibody Tests / Kit: 363 (H917)

Required Calibrator:

KAI-016C	**Multi-Analyte Calibrator** 6 x 1 mL, liquid For 6 point calibration

Recommended Control:

K49C-4M	**Liquid ITA Control** L1: 2 x 2 mL, liquid L2: 2 x 2 mL, liquid

Immunoassay Reagents for Chemistry Analyzers

Immunoglobulin G

Method: Immunoturbidimetric
Format: Liquid
Linearity: 100 - 4,500 mg/dL
Sample: Serum

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-014	**IgG** R1: 4 x 20 mL buffer R2: 4 x 20 mL antibody Tests / Kit: 444 (H917)

Required Calibrator:

KAI-016C	**Multi-Analyte Calibrator** 6 x 1 mL, liquid For 6 point calibration

Recommended Control:

K49C-4M	**Liquid ITA Control** L1: 2 x 2 mL, liquid L2: 2 x 2 mL, liquid

Immunoglobulin M

Method: Immunoturbidimetric
Format: Liquid
Linearity: 10 - 350 mg/dL
Sample: Serum

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-015	**IgM** R1: 4 x 20 mL buffer R2: 2 x 10 mL antibody Tests / Kit: 363 (H917)

Required Calibrator:

KAI-016C	**Multi-Analyte Calibrator** 6 x 1 mL, liquid For 6 point calibration

Recommended Control:

K49C-4M	**Liquid ITA Control** L1: 2 x 2 mL, liquid L2: 2 x 2 mL, liquid

Insulin

Method: Immunoturbidimetric
Format: Liquid
Linearity: 1 - 100 µIU/mL
Sample: Serum, Plasma

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-071	**Insulin** R1: 2 x 13.5 mL buffer R2: 2 x 5 mL antibody Tests / Kit: 200 (H917)

Required Calibrator:

KAI-072C	**Insulin Calibrator** 5 x 0.5 mL, lyophilized For 6 point calibration

Recommended Control:

K73C-4M	**Insulin Control** L1: 2 x 0.5 mL, lyophilized L2: 2 x 0.5 mL, lyophilized

Immunoassay Reagents for Chemistry Analyzers

KL-6 (Ver.2)
Krebs von den Lungen-6

Method: Immunoturbidimetric
Format: Liquid
Linearity: 70 - 10,000 U/mL
Sample: Serum, Plasma

For Research Use Only in the U.S. Not for use in diagnostic procedures in the U.S.

NEW PRODUCT

Cat. No.	Description
KAI-350*	**KL-6 (Ver.2)*** R1: 1 x 15 mL buffer R2: 1 x 5 mL antibody Tests / Kit: 100 (H917)

Required Calibrator:

KAI-351C*	**KL-6 Calibrator (Ver.2)*** 5 x 1 mL, liquid For 5 point calibration

Recommended Control:

K352C-2M*	**KL-6 Control (Ver.2)*** L1: 1 x 2 mL, liquid L2: 1 x 2 mL, liquid

Lipoprotein(a)

Method: Immunoturbidimetric
Format: Liquid
Linearity: 5 - 150 mg/dL
Sample: Serum

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-044	**Lp(a)** R1: 1 x 21 mL buffer R2: 1 x 3 mL antibody Tests / Kit: 85 (H917)

Required Calibrator:

KAI-018C	**Lp(a) Calibrator** 1 x 1 mL, lyophilized 2 x 2 mL liquid diluent For 6 point calibration

Recommended Control:

K114C-4M	**Lp(a) Control** L1: 2 x 1 mL, lyophilized L2: 2 x 1 mL, lyophilized

Microalbumin

Method: Immunoturbidimetric
Format: Liquid
Linearity: 0.2 - 30 mg/dL
Sample: Urine

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-019	**Microalbumin** R1: 4 x 20 mL buffer R2: 2 x 10 mL antibody Tests / Kit: 285 (H917)

Cat. No.	Description
KAI-057	**Microalbumin (L)** R1: 3 x 80 mL buffer R2: 1 x 80 mL antibody Tests / Kit: 1142 (H917)

Required Calibrator:

KAI-020C	**Microalbumin Calibrator** 6 x 1 mL, liquid For 6 point calibration

Recommended Control:

K37C-4M	**Microalbumin Urine Control** L1: 2 x 7 mL, liquid L2: 2 x 7 mL, liquid

* For research use only in the U.S. Not for use in diagnostic procedures in the U.S.

Immunoassay Reagents for Chemistry Analyzers

Prealbumin

Method: Immunoturbidimetric
Format: Liquid
Linearity: 0 - 60 mg/dL
Sample: Serum

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-053	**Prealbumin** R1: 4 x 20 mL buffer R2: 2 x 8 mL antibody Tests / Kit: 400 (H917)

Required Calibrator:

KAI-054C	**Prealbumin Calibrator** 6 x 1 mL, liquid For 6 point calibration

Recommended Control:

K49C-4M	**Liquid ITA Control** L1: 2 x 2 mL, liquid L2: 2 x 2 mL, liquid

Remnant Lipoprotein Cholest.

Method: Colorimetric
Format: Liquid
Linearity: 1 - 80 mg/dL
Sample: Serum

For Research Use Only in the U.S. Not for use in diagnostic procedures in the U.S.

Cat. No.	Description
KAI-260*	**RLP-C*** R1: 1 x 12 mL buffer R2: 1 x 4 mL enzyme Tests / Kit: 80 (H917)

Required Calibrator:

KAI-261C*	**RLP-C Calibrator*** 1 x 1 mL, lyophilized For 2 point calibration

Recommended Control:

K262C-6M*	**RLP-C Control*** L1: 3 x 1 mL, lyophilized L2: 3 x 1 mL, lyophilized

Rheumatoid Factor (Ver.2)

Method: Immunoturbidimetric
Format: Liquid
Linearity: 6.65 - 600 IU/mL
Sample: Serum

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-230	**RF (Ver.2)** R1: 1 x 20 mL buffer R2: 1 x 20 mL IgG Tests / Kit: 320 (H917)

Required Calibrator:

KAI-231C	**RF Calibrator** 6 x 1 mL, liquid For 6 point calibration

Recommended Control:

K55C-4M	**ASO/RF/CRP Control** L1: 2 x 2 mL, lyophilized L2: 2 x 2 mL, lyophilized

* For research use only in the U.S. Not for use in diagnostic procedures in the U.S.

Immunoassay Reagents for Chemistry Analyzers

Transferrin

Method: Immunoturbidimetric
Format: Liquid
Linearity: 45 - 455 mg/dL
Sample: Serum

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-023	**Transferrin** R1: 3 x 20 mL buffer R2: 2 x 10 mL antibody Tests / Kit: 285 (H917)

Required Calibrator:

KAI-016C	**Multi-Analyte Calibrator** 6 x 1 mL, liquid For 6 point calibration

Recommended Control:

K49C-4M	**Liquid ITA Control** L1: 2 x 2 mL, liquid L2: 2 x 2 mL, liquid

Unsaturated Iron Binding Capacity

Method: Colorimetric
Format: Liquid
Linearity: 14 - 700 µg/dL
Sample: Serum

For *In Vitro* Diagnostic Use.

Cat. No.	Description
KAI-300	**UIBC** R1: 1 x 40 mL Ferric Ammonium Sulfate R2: 1 x 20 mL Nitroso-PSAP Tests / Kit: 200 (H917)

Required Calibrator:

KAI-302C	**UIBC Calibrator** 1 x 5 mL, liquid For 2 point calibration

Calibrators

Cat. No.	Description
KAI-068C	**Serum/Plasma Alpha-1 Microglobulin Calibrator** 5 x 1 mL, liquid For 6 point calibration

Cat. No.	Description
KAI-067C	**Urine Alpha-1 Microglobulin Calibrator** 4 x 1 mL, liquid For 5 point calibration

Cat. No.	Description
KAI-079C	**ASO Calibrator** 5 x 1 mL, liquid For 5 point calibration

Cat. No.	Description
KAI-008C	**Apo AI/B Calibrator** 1 x 1 mL, lyophilized 2 x 2 mL liquid diluent For 6 point calibration

Cat. No.	Description
KAI-041C*	**Apo AII/CII/CIII Calibrator*** 1 x 1 mL, lyophilized 2 x 2 mL liquid diluent For 6 point calibration

Cat. No.	Description
KAI-025C*	**Apo E Calibrator*** 1 x 1 mL, lyophilized 2 x 2 mL liquid diluent For 6 point calibration

Cat. No.	Description
KAI-281C	**B2MG Serum / Plasma Calib.** 5 x 2 mL, liquid For 6 point calibration

Cat. No.	Description
KAI-282C	**B2MG Urine Calibrator** 5 x 2 mL, liquid For 6 point calibration

Cat. No.	Description
KAI-012C	**CRP Calibrator** 6 x 1 mL, liquid For 6 point calibration

* For research use only in the U.S. Not for use in diagnostic procedures in the U.S.

Calibrators

Cat. No.	Description
KAI-161C	**hsCRP Calibrator** 4 x 1 mL, liquid For 5 point calibration

Cat. No.	Description
KAI-084C	**CRP Calibrator E** 5 x 2 mL, liquid For 6 point calibration with linear range: 0.05 – 160 mg/L

Cat. No.	Description
KAI-099C	**Cystatin C Calibrator** 6 x 1 mL, liquid For 6 point calibration

Cat. No.	Description
KAI-091C	**D-Dimer Calibrator** 6 x 1 mL, lyophilized For 6 point calibration

Cat. No.	Description
KAI-103C**	**High-Sensitive** **D-Dimer Calibrator**** 5 x lyophilized stock 1 x 40 mL liquid diluent For 6 point calibration

Cat. No.	Description
KAI-106C	**Factor XIII Calibrator** 5 x 1.0 mL, lyophilized 1 x 40 mL liquid diluent For 6 point calibration

Cat. No.	Description
KAI-112C*	**Plasma FDP Calibrator*** 5 x 1.0 mL, lyophilized 1 x 40 mL liquid diluent For 6 point calibration

Cat. No.	Description
KAI-321C*	**Serum FDP Calibrator*** 5 x 1.0 mL, lyophilized 1 x 40 mL liquid diluent For 6 point calibration

Cat. No.	Description
KAI-326C*	**Urine FDP Calibrator*** 5 x 1.0 mL, lyophilized 1 x 40 mL liquid diluent For 6 point calibration

* For research use only in the U.S. Not for use in diagnostic procedures in the U.S.
** Not for use in the U.S.

Calibrators

Cat. No. **KAI-094C**

Description
Ferritin Calibrator
4 x 1 mL, liquid

For 5 point calibration

Cat. No. **KAI-136C**

Description
Fibrinogen Calibrator
1 x 1 mL, lyophilized

For 4 point calibration

Cat. No. **KAI-241C***

Description
H. pylori Calibrator*
5 x 2 mL, liquid

For 6 point calibration

Cat. No. **KAI-098C**

Description
Hemoglobin A1c Calibrator
4 x 0.5 mL, lyophilized

For 5 point calibration

Cat. No. **KAI-072C**

Description
Insulin Calibrator
5 x 0.5 mL, lyophilized

For 6 point calibration

Cat. No. **KAI-351C***

Description
KL-6 Calibrator (Ver.2)*
5 x 2 mL, liquid

For 5 point calibration

Cat. No. **KAI-018C**

Description
Lp(a) Calibrator
1 x 1 mL, lyophilized
2 x 2 mL liquid diluent

For 6 point calibration

Cat. No. **KAI-020C**

Description
Microalbumin Calibrator
6 x 1 mL, liquid

For 6 point calibration

Cat. No. **KAI-016C**

Description
Multi-Analyte Calibrator
6 x 1 mL, liquid
For 6 point calibration
For calibration of: Complement C3 and C4, Haptoglobin, IgA, IgG, IgM, Transferrin

* For research use only in the U.S. Not for use in diagnostic procedures in the U.S.

Calibrators

Cat. No.	Description
KAI-054C	**Prealbumin Calibrator** 6 x 1 mL, liquid For 6 point calibration



Cat. No.	Description
KAI-261C*	**RLP-C Calibrator*** 1 x 1 mL, lyophilized For 2 point calibration

Cat. No.	Description
KAI-231C	**RF Calibrator (Ver.2)** 6 x 1 mL, liquid For 6 point calibration

Cat. No.	Description
KAI-302C	**UIBC Calibrator** 1 x 5 mL, liquid For 2 point calibration

Cat. No.	Description
KAI-109D*	**Urine FDP Sample Diluent*** 1 x 40 mL liquid diluent For dilution of high concentration samples

* For research use only in the U.S. Not for use in diagnostic procedures in the U.S.

Controls

NEW PRODUCT

Cat. No.	Description
K310C-4M*	**Apo AI / B Control***
	L1: 2 x 1 mL, lyophilized
	L2: 2 x 1 mL, lyophilized

Assayed for: Apolipoprotein AI and B

NEW PRODUCT

Cat. No.	Description
K315C-4M*	**Apo AII / CII / CIII / E Control***
	L1: 2 x 1 mL, lyophilized
	L2: 2 x 1 mL, lyophilized

Assayed for: Apolipoprotein AII, CII, CIII, and E

Cat. No.	Description
K55C-4M	**ASO/RF/CRP Control**
	L1: 2 x 2 mL, lyophilized
	L2: 2 x 2 mL, lyophilized

Assayed for: Anti-Streptolysin O, Rheumatoid Factor, C-Reactive Protein

Cat. No.	Description
K283C-4M	**B2MG / Ferritin - Ser / Pl Control**
	L1: 2 x 3 mL, lyophilized
	L2: 2 x 3 mL, lyophilized

Assayed for: Beta-2 Microglobulin, Ferritin

Cat. No.	Description
K284C-2M	**B2MG Urine Control**
	L1: 1 x 2 mL, liquid
	L2: 1 x 2 mL, liquid

Assayed for: Beta-2 Microglobulin

Cat. No.	Description
K180C-2M	**hsCRP Control**
	L1: 1 x 3 mL, liquid
	L2: 1 x 3 mL, liquid

Assayed for: CRP

Cat. No.	Description
K100C-4M	**Cystatin C Control**
	L1: 2 x 1 mL, liquid
	L2: 2 x 1 mL, liquid

Assayed for: Cystatin C

Cat. No.	Description
K92C-4M	**D-Dimer Control**
	L1: 2 x 1 mL, lyophilized
	L2: 2 x 1 mL, lyophilized

Assayed for: Beta-2 Microglobulin, Ferritin

Cat. No.	Description
K115C-10M**	**High-Sensitive D-Dimer Control****
	L1: 5 x 0.5 mL, lyophilized
	L2: 5 x 0.5 mL, lyophilized

Assayed for: High-Sensitive D-Dimer

* For research use only in the U.S. Not for use in diagnostic procedures in the U.S.

** Not for use in the U.S.

Controls

Cat. No. **K135C-10M**

Description
Factor XIII Control
L1: 5 x 0.5 mL, lyophilized
L2: 5 x 0.5 mL, lyophilized

Assayed for: Factor XIII

Cat. No. **K131C-10M***

Description
Plasma FDP Control*
L1: 5 x 0.5 mL, lyophilized
L2: 5 x 0.5 mL, lyophilized

Assayed for: Plasma FDP

Cat. No. **K322C-10M***

Description
Serum FDP Control*
L1: 5 x 0.5 mL, lyophilized
L2: 5 x 0.5 mL, lyophilized

Assayed for: Serum FDP

NEW PRODUCT

Cat. No. **K327C-10M***

Description
Urine FDP Control*
L1: 5 x 0.5 mL, lyophilized
L2: 5 x 0.5 mL, lyophilized

Assayed for: Urine FDP

Cat. No. **K63C-10M**

Description
Fibrinogen Control
L1: 5 x 1 mL, lyophilized
L2: 5 x 1 mL, lyophilized

Assayed for: Fibrinogen (Antigen)

Cat. No. **K242C-2M***

Description
H. pylori Control*
L1: 1 x 2 mL, liquid
L2: 1 x 2 mL, liquid

Assayed for: H. pylori

Cat. No. **K29C-4M**

Description
Hemoglobin A1c Control
L1: 2 x 0.5 mL, lyophilized
L2: 2 x 0.5 mL, lyophilized

Assayed for: Hemoglobin A1c

Cat. No. **K94C-4M**

Description
Total IgE Control
L1: 2 x 1 mL, liquid
L2: 2 x 1 mL, liquid

Assayed for:Total IgE

Cat. No. **K73C-4M**

Description
Insulin Control
L1: 2 x 0.5 mL, lyophilized
L2: 2 x 0.5 mL, lyophilized

Assayed for: Insulin

* For research use only in the U.S. Not for use in diagnostic procedures in the U.S.

Controls

Cat. No.	Description
K352C-2M*	**KL-6 Control (Ver.2)***
	L1: 1 x 2 mL, liquid
	L2: 1 x 2 mL, liquid

Assayed for: KL-6

Cat. No.	Description
K49C-4M	**Liquid ITA Control**
	L1: 2 x 2 mL, liquid
	L2: 2 x 2 mL, liquid

Assayed for: Complement C3 and C4, CRP, Haptoglobin, IgA, IgG, IgM, Prealbumin, Transferrin

Cat. No.	Description
K114C-4M	**Lp(a) Control**
	L1: 2 x 1 mL, lyophilized
	L2: 2 x 1 mL, lyophilized

Assayed for: Lipoprotein(a)

Cat. No.	Description
K37C-4M	**Microalbumin Urine Control**
	L1: 2 x 7 mL, liquid
	L2: 2 x 7 mL, liquid

Assayed for: Creatinine, Microalbumin

Cat. No.	Description
K262C-6M*	**RLP-C Control***
	L1: 3 x 1 mL, lyophilized
	L2: 3 x 1 mL, lyophilized

Assayed for: RLP-C

* For research use only in the U.S. Not for use in diagnostic procedures in the U.S.

Available Chemistry Analyzer Applications

Preliminary applications may be available for unmarked analyzer/assay combinations.
Please contact us for more information.

	Abbott Aeroset	Abbott c8000/c16000/ci16200	Alfa Wassermann ACE/Alera/Axcel	Beckman/Olympus Reply/AU560	Beckman/Olympus AU400/AU480	Beckman/Olympus AU600/640/680	Beckman/Olympus AU800	Beckman/Olympus AU1000	Beckman/Olympus AU2700/5400	Beckman/Olympus AU5000	Beckman Synchron CX	Beckman Synchron DxC/LX	Horiba ABX Pentra 400	ILAB 600/900/1800	Mindray BS-200	Roche Cobas c501/6000	Roche Cobas Fara	Roche Cobas Mira	Roche/Hitachi 704	Roche/Hitachi 717	Roche/Hitachi 736	Roche/Hitachi 747	Roche/Hitachi 902	Roche/Hitachi 911	Roche/Hitachi 912	Roche/Hitachi 914	Roche/Hitachi 917	Roche/Hitachi Modular P	Siemens/Bayer Advia 1650	Siemens/Bayer Advia 2400	Siemens/Bayer AXON/60R	Siemens/Dade Dimension	Stanbio Sirrus/Prestige 24i	Vital Diagnostics Envoy 500
UIBC	●	●	●		●	●			●			●	●		❍	●				●		●	●	●	●		●	●	●	●		●	●	
Transferrin	●	●	●	●	●	●	●	●	●		●	●	●	●	❍	❍	●	●	●	●	●	●	●	●	●	●	●	●	●	●		●	●	❍
RLP-C	●	●	❍		❍	❍			❍				❍		❍	❍		❍						●	●		●	●	●	●		❍	❍	❍
RF (Ver.2)	●	●		●	●	●	●	●	●	●	●	●	●	●	❍	❍	●	●	●	●	●	●	●	●	●	●	●	●	●	●		●	●	●
Prealbumin	●	●	●		●	●	●	●	●		●	●	❍	●	❍	❍	●	●	●	●		●	●	●	●	●	●	●	●	●		●	●	❍
Microalbumin	●	●	●	●	●	●	●	●	●	●	●	●	❍	●	❍	❍	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	❍
Lp(a)	●	●	●	●	●	●	●	●	●		●	●	❍	●	❍	❍	●	●	●	●	●	●	●	●	●	●	●	●	●	●		●	●	❍
KL-6 (Ver.2)	●	●	❍		❍	❍			❍				❍		❍	❍		❍						❍	❍		●	●	❍	●		❍	❍	❍
Insulin	❍	❍			●	●			●		●	●	❍		❍	❍	❍	●	●	●		❍	❍	●	●		●	●	●	●		❍	●	❍
IgM	●	●	●	●	●	●	●	●	●		●	●	●	●	❍	❍	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●
IgG	●	●	●	●	●	●	●	●	●		X	X	●	●	❍	❍	X	X	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●
IgA	●	●	●	●	●	●	●	●	●		●	●	●	●	❍	❍	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●
HbA1c	❍	●			●	●			●		●	❍	❍		●	❍		●	●	●		❍	●	●	❍	●	●	●	●			●	●	❍
Haptoglobin	●	●	●	●	●	●	●	●	●		●	●	❍	●	❍	❍	●	●	●	●		●	●	●	●	●	●	●	●	●		●	●	❍
H. pylori	❍	●	❍		●	●			●			❍	❍		❍	●		❍						●	❍		●	●	❍	❍		❍	●	❍
Fibrinogen	●	●	●	●	●	●	●	●	●		●	●	❍	●	❍	❍	●	●	●	●		●	●	●	●	●	●	●	●	●		●	●	❍
Ferritin	●	●	●		●	●			●		●	●	●		❍	❍		●	●	●		●	●	●	●		●	●	●	●		●	❍	❍
D-Dimer	●	●			●	●			●		●	●	❍		❍	❍		●	●	●		❍	❍	●	❍		●	●	●	●		❍		❍
Cystatin C	●	●			●	●			●														●	●			●	●	●	●			❍	
CRP (3) + Cal. E	●	●			●	●		●	●		X	❍	●		❍	❍	❍	●		●		●	●	●	●		●	●	●	●		●	●	❍
hsCRP		●	❍		●	●			●		●	❍	●		●	●		X		❍				●			●		❍	❍		❍	●	❍
CRP	●	●	●	●	●	●	●	●	●	●	●	●	❍	●	❍	❍	●	●	●	●		●	●	●	●	●	●	●	●	●	●	●	●	●
Complement C4	●	●	●	●	●	●	●	●	●		●	●	●	●	❍	❍	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●
Complement C3	●	●	●	●	●	●	●	●	●		●	●	●	●	❍	❍	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●
Beta-2 Microglobulin		●			●	●			●							●								●			●	●	●	●			●	
Apo E	●	●	●	●	●	●	●	●	●	●	●	●	❍	●	❍	❍	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	❍
Apo CIII	●	●			●	●	●	●	●		●	●	❍	●	❍	●	●	●	●	●		●	●	●	●	●	●	●	●	●		●	●	❍
Apo CII	●	●	X		●	●	●	●	●		●	●	❍	●	❍	❍	●	●	●	●		●	●	●	●	●	●	●	●	●		●	●	❍
Apo B	●	●	●	●	●	●	●	●	●	●	●	●	❍	●	❍	❍	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	❍
Apo AII	●	●	●		●	●	●	●	●	●	●	●	❍	●	❍	❍	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	❍
Apo AI	●	●	●	●	●	●	●	●	●	●	●	●	●	●	❍	❍	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	●	❍
ASO (WHO)	●	●			●	●			●		X	X	❍		❍	❍	X	X	●	●				●	●	●	●	●	●	●		●	●	❍
α-1 Microglobulin	●	●			●	●			●		❍		❍		❍	❍		●		●		❍	●	●	●		●	●	❍	❍			●	❍

Key: ● = Application Available ❍ = Preliminary Application Available X = Not Adaptable

800-526-4925 / 206-575-8068 • FAX: 206-575-8094
www.k-assay.com

Approximate Number of Tests per Kit

(On selected chemistry analyzers)

	Abbott Aeroset	Abbott c8000/c16000/ci16200	Alfa Wassermann ACE/Alera/Axcel	Beckman/Olympus Reply/AU560	Beckman/Olympus AU400/AU480	Beckman/Olympus AU600/640/680	Beckman/Olympus AU800	Beckman/Olympus AU1000	Beckman/Olympus AU2700/5400	Beckman/Olympus AU5000	Beckman Synchron CX	Beckman Synchron DxC/LX	Horiba ABX Pentra 400	ILAB 600/900/1800	Mindray BS-200	Roche Cobas c501/6000	Roche Cobas Fara	Roche Cobas Mira	Roche/Hitachi 704	Roche/Hitachi 717	Roche/Hitachi 736	Roche/Hitachi 747	Roche/Hitachi 902	Roche/Hitachi 911	Roche/Hitachi 912	Roche/Hitachi 914	Roche/Hitachi 917	Roche/Hitachi Modular P	Siemens/Bayer Advia 1650	Siemens/Bayer Advia 2400	Siemens/Bayer AXON/60R	Siemens/Dade Dimension	Stanbio Sirrus/Prestige 24i	Vital Diagnostics Envoy 500
UIBC	250	333	200		250	250			250			200	250		❍	333				200		200	200	200	200		200	250	285	666		200	200	
Transferrin	285	285	200	200	285	285	240	285	285		266	285	235	214	❍	❍	222	222	171	200	200	200	285	200	285	200	285	285	425	600		181	200	❍
RLP-C	100	66 / 100	❍		❍	❍			❍				❍		❍	❍		❍						40	80		66	80	120	120		❍	❍	❍
RF (Ver.2)	200	200		100	153	153	100	153	153	100	160	160	200	133	❍	❍	200	200	142	133	133	133	133	133	166	133	166	200	250	400		100	200	133
Prealbumin	533	500	320		533	533	320	533	533		320	320	❍	266	❍	❍	320	320	228	320		320	320	320	533	320	400	533	571	800		228	400	❍
Microalbumin (L)	1600	1500	800	800	1600	1600	800	1600	1600	800	1142	1142	❍	800	❍	❍	888	888	666	800	800	800	1142	800	1600	800	1142	1600	1702	2400	800	727	1142	❍
Microalbumin	400	377	200	200	400	400	200	400	400	200	285	285	❍	200	❍	❍	222	222	166	200	200	200	285	200	400	200	285	400	425	606	200	181	285	❍
Lp(a)	136	125	60	75	120	120	60	120	120		85	85	❍	60	❍	❍	75	75	60	60	60	60	85	60	136	60	85	136	150	210		60	85	❍
KL-6 (Ver.2)	166	166	❍		❍	❍			❍				❍		❍	❍		❍						❍	❍		100	100	❍	250		❍	❍	❍
Insulin	❍	❍			200	200			200		184	184	❍		❍	❍	❍	153	142	142		❍	❍	153	153		200	200	400	450		❍	142	❍
IgM	444	363	285	285	444	444	285	444	444		285	357	285	200	❍	❍	285	285	200	285	200	285	363	285	444	285	363	444	500	714	285	200	285	285
IgG	533	500	320	320	500	500	320	533	500		X	X	363	285	❍	❍	X	X	228	320	228	320	444	320	533	320	444	533	571	800	320	235	400	320
IgA	444	363	285	285	444	444	285	444	444		285	357	285	200	❍	❍	285	285	200	285	200	285	363	285	444	285	363	444	500	714	285	200	285	285
HbA1c (L)	❍	600			600	600			600		447	❍	❍		500	❍		600	300	500		❍	500	473	❍	500	600	600	1200			447	428	❍
HbA1c	❍	200			200	200			200		149	❍	❍		166	❍		200	100	166		❍	166	157	❍	166	200	200	400			149	142	❍
Haptoglobin	311	291	200	200	311	311	200	311	311		200	200	❍	140	❍	❍	200	200	140	200		200	254	200	311	140	254	311	358	500		140	200	❍
H. pylori	❍	300	❍		260	260			260			❍	❍		❍	400		❍						206	❍		266	266	❍	❍		❍	300	❍
Fibrinogen	266	250	160	160	266	266	160	266	266		160	160	❍	125	❍	❍	160	160	114	160		160	160	160	266	160	160	266	285	400		114	200	❍
Ferritin	225	150	105		112	112			112		128	128	105		❍	❍		105	75	100		90	100	100	128		128	128	225	300		90	❍	❍
D-Dimer	133	133			160	160			160		113	113	❍		❍	❍		114	80	80		❍	❍	80	❍		114	114	266	266		❍		❍
Cystatin C	200	200			166	166			166		X	X						X					166	166			200	200	333	400			❍	
CRP (3) + Cal. E	333	333			333	333		277	333		X	❍	526		❍	❍	❍	500		333		333	333	333	333		333	333	625	625		333	333	❍
hsCRP		200	❍		250	250			250		200	❍	400		200	277		X		❍				200			200		❍	❍		❍	200	❍
CRP	500	500	285	285	500	500	285	500	500	285	333	333	❍	285	❍	❍	285	285	200	285		285	333	285	500	285	333	500	571	800	285	250	400	285
Complement C4	444	416	285	285	444	444	285	444	444		285	285	285	200	❍	❍	285	285	171	285	200	285	363	285	444	285	363	444	512	714	285	200	285	285
Complement C3	444	416	285	285	444	444	285	444	444		285	285	285	200	❍	❍	285	285	171	285	200	285	363	285	444	285	363	444	512	714	285	200	285	285
Beta-2 Microglobulin		333U, 250S/P			250	250			250							333U, 500S/P								175U, 233S/P			250U, 333S/P	250U, 333S/P	625	625			166U, 200S/P	
Apo E	285	285	200	200	285	285	200	285	285	200	266	266	❍	200	❍	❍	235	235	166	200	200	200	285	200	285	200	285	285	425	600	200	181	285	❍
Apo CIII	285	285			285	285	200	285	285		266	266	❍	200	❍	400	214	214	166	200		200	285	200	285	200	285	285	425	600		181	285	❍
Apo CII	285	285	X		285	285	200	285	285		266	266	❍	200	❍	❍	250	250	166	200		200	285	200	285	200	285	285	425	600		181	285	❍
Apo B	285	285	200	200	285	285	200	285	285	200	266	266	❍	200	❍	❍	250	250	166	200	200	200	285	200	285	200	285	285	425	600	200	181	285	❍
Apo AII	285	285	200		285	285	200	285	285	200	250	250	❍	200	❍	❍	222	222	200	200	200	200	285	200	285	200	285	285	500	714	200	200	250	❍
Apo AI	285	285	200	200	285	285	200	285	285	200	266	266	235	200	❍	❍	235	235	166	200	200	200	285	200	285	200	285	285	425	600	200	181	285	❍
ASO (WHO)	571	571			571	571	400	571	571		X	X	❍		❍	❍	X	X	400	400		400	571	400	571	400	571	571	800	1000		400	400	❍
α-1 Microglobulin	333	333			277	277			277		❍		❍		❍	❍		416U, 500S/P		400U, 277S/P		❍	312U, 250S/P	312U, 250S/P	500U, 357S/P		500U, 357S/P	500U, 357S/P	❍	❍			277	❍

Key: ❍ = Preliminary Application Available X = Not Adaptable

800-526-4925 / 206-575-8068 • FAX: 206-575-8094
www.k-assay.com

ORDERING / PRICING / TECHNICAL INFORMATION:

Telephone, mail, e-mail, or FAX to :

KAMIYA BIOMEDICAL COMPANY
12779 Gateway Drive S
Tukwila, WA 98168 USA

206-575-8068 / 800-526-4925 / 800-KAMIYA-5 • FAX: 206-575-8094
www.k-assay.com diagnostics@k-assay.com

Please include the following information with all orders:

1. **Name, Telephone Number, FAX Number, Shipping Address, and Billing Address.**
2. **Quantity, Product Name, and Catalog Number.**
3. **Purchase Order Number.**

PRICES: F.O.B. KAMIYA - Tukwila, WA. Prices are subject to change without notice. Additional charges for handling, freight, and taxes will be added, as applicable.

TERMS: U.S. Customers: Net 30 days from date of invoice. Overdue amounts are subject to 1.5% monthly late charge. Prepayments may be required for certain customers.
Prices are subject to change without notice.

International Customers: Prepayment in full is required before shipment.

FREIGHT: Prepaid and added to invoice.

RETURNS: Due to the perishable nature of the products, no returns will be allowed. Please check your order carefully before submitting to us.

TECHNICAL SERVICE:

Our website (www.k-assay.com) contains our full catalog and technical information. For additional technical service, please contact us by e-mail (diagnostics@k-assay.com), phone, or FAX.

CONDITIONS OF SALE:

All products labeled “For Research Use” must be used for research use only in the U.S. and not for *in vitro* diagnostic or clinical use. All products should be used in accordance with their labeled intended use.

KAMIYA warrants that its products conform to the information contained in this and other KAMIYA publications. Website information may not be up to date. Purchaser must determine suitability of the product for his/her particular use. THIS WARRANTY IS IN LIEU OF ALL OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED, INCLUDING WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE. IN NO CASE SHALL KAMIYA BE LIABLE FOR LOST PROFITS OR DIRECT, INDIRECT, INCIDENTAL, OR CONSEQUENTIAL DAMAGES INCURRED BY THE BUYER OR THIRD PARTIES. By accepting shipments from KAMIYA, buyer explicitly recognizes and agrees to these Conditions of Sale.

Please inspect all products immediately upon receipt. Any claims for damaged, leaking, or missing products must be reported in writing to KAMIYA BIOMEDICAL COMPANY by the buyer within three (3) days from the date of receipt of the products. No replacement or refund will be provided to the customer after 3 days for damaged, leaking, or missing products.

K-ASSAY® *The Assay You Can Trust . . .*



KAMIYA BIOMEDICAL COMPANY

12779 Gateway Drive S
Tukwila, WA 98168 USA

206-575-8068 / 800-526-4925 / 800-KAMIYA-5 • FAX: 206-575-8094

www.k-assay.com **diagnostics@k-assay.com**